


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 26223 Accredited to ISO/IEC 17025:2017	<b>Storm Aviation Limited</b>	
	Issue No: 003    Issue date: 21 February 2025	
	<b>Storm Aviation Limited</b> Diamond Hangar Long Border Road Stansted CM24 1RE	<b>Contact: Michael Reed</b> Tel: +44 (0)1279 669014 E-Mail: <a href="mailto:calibration@stormaviation.com">calibration@stormaviation.com</a> Website: <a href="http://www.stormaviation.com">www.stormaviation.com</a>
<b>Calibration performed at the above address only</b>		

### Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks
<b>PRESSURE</b>  Hydraulic pressure (gauge)  Calibration of pressure indicating instruments and gauges	344 kPa to 138 MPa	Q [ 0.012 %, 345 Pa ]	Methods consistent with EURAMET CG17
<b>TORQUE</b>  Torque Wrenches	0.4 N·m to 2 N·m 2 N·m to 50 N·m 10 N·m to 250 N·m 40 N·m to 1000 N·m	1.2 % 1.0 % 0.50 % 0.55 %	BS EN ISO 6789-2:2017  1. The uncertainty quoted is for both the application of the calibration torque and the characteristics of the device being calibrated  2. Calibrations may be given in other torque units lbf.in, lbf.ft, oz.in.
END			



26223

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Storm Aviation Limited**  
**Issue No: 003 Issue date: 21 February 2025**

Calibration performed at main address only

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

**Expression of CMCs - symbols and units**

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$